



EMSL Analytical, Inc.

1101-A Aviation Parkway Morrisville, NC 27560

Phone: (919) 465-3900

Fax: (919) 465-3900

Email: raleighlab@emsl.com

Attn: Jeff Stockweather
 Mold Solutions, Inc.
 4458 Augusta Road
 Suite 4-C
 Lexington, SC 29073

Proj: B. Bledsoe, 41 Boardwalk

EMSL Order: 290801247
 Customer ID: MLDS78
 Received: 3/5/08
 Analyzed: 3/6/08
 Report Date: 3/6/08

Air-O-Cell™ Cassette Analysis of Fungal Spores & Other Airborne Particulates by Optical Microscopy (EMSL Method M001)

Lab Sample Number:	290801247-0001	290801247-0002	290801247-0003		
Client Sample ID:	13320270	13320216	13320198		
Volume (L):	75	75	75		
Sample Location:	Living Room	Mstr Bedroom	Outside		
Spore Types	Count/m ³	Count/m ³	Count/m ³		
Agrocybe/Coprinus	-	-	-		
Alternaria	-	42	-		
Ascospores	-	126	42		
Aspergillus/Penicillium	966	-	-		
Basidiospores	-	126	210		
Bipolaris	-	42	-		
Chaetomium	-	-	-		
Cladosporium	714	168	210		
Curvularia	-	42	-		
Epicoccum	126	-	294		
Fusarium	-	-	-		
Ganoderma	-	-	-		
Myxomycete	42	-	-		
Paecilomyces	-	-	-		
Rust	-	-	-		
Scopulariopsis	-	-	-		
Stachybotrys	-	-	-		
Torula	-	-	-		
Ulocladium	-	-	-		
Unidentifiable Spores	84	-	42		
Zygomycetes	-	-	-		
Nigrospora	84	84	-		
Pithomyces	-	42	42		
Total Fungi	2020	672	840		
Fibrous Particulate	126	672	-		
Hyphal Fragment	546	294	1510		
Insect Fragment	-	42	42		
Pollen	13*	42	252		
Analyt. Sensitivity 600x	42	42	42		
Analyt. Sensitivity 300x	13*	13*	13*		
Skin Fragments (1-4)	3	3	1		
Background (1-5)	2	2	1		

No discernable field blank was submitted with this group of samples.

Accreditations: AIHA EMLAP 173741

High levels of background particulate can obscure spores and other particulates leading to underestimation. Background levels of 5 indicate an overloading of background particulates, prohibiting accurate detection and quantification. Present = Spores detected on overloaded samples. The detection limit is equal to one fungal spore, structure, pollen, fiber particle or insect fragment. "*** Denotes particles found at 300X. This report relates only to the samples reported and may not be reproduced, except in full, without written approval by EMSL. Results have not been adjusted for field or laboratory blank unless otherwise noted. Samples received in good condition unless otherwise noted.

sdv2.7.7.8

Alan Goldstein, Ph.D., Laboratory Manager
 or Other Approved Signatory